

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients
Dr. Matthew Arduino, CDC
A Webber Training Teleclass

Preventing Adverse Outcomes In Maintenance Hemodialysis Patients

Matthew J. Arduino, Dr.P.H.
National Center for Preparedness, Detection,
and Control of Infectious Disease
Centers for Disease Control and Prevention
matthew.arduino@cdc.hhs.gov


Hosted by Paul Webber
paul@webbertraining.com

The findings and conclusions in this presentation are those of the author and do not necessarily represent the views of the Centers for Disease Control and Prevention

Adverse Outcomes in ESRD

- Insufficient treatment
- Dialysis-Associated Infections
- Intoxications/Chemical Poisoning
 - ◆ Water treatment failures
 - ◆ System design
 - ◆ Human error
 - ◆ Manufacturing error
- Allergic Reactions
- Non-chemical associated hemolytic events
- Bleeding

Hemodialysis



I. Intoxications/Chemical Poisonings


- Failure to rinse germicide from dialyzers or hemodialysis system
- Water Treatment Issues
 - ◆ Trace elements in water
 - ◆ Malfunction of water treatment device
 - ◆ Biologic Toxins in water supply
- Dialysate Quality
 - ◆ Use of acid concentrate instead of acetate concentrate

Hemodialysis

- Hemodialysis patients are exposed to approximately 300-600 Liters of water/week.
- On average approximately 16-24 L of water is ingested per week.

Feed Water

- Dialysis centers use water from a public/private supply, which may be derived from either surface or ground waters



Hosted by Paul Webber paul@webbertraining.com
www.webbertraining.com

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients
Dr. Matthew Arduino, CDC
A Webber Training Teleclass

Water Contaminants

- **Biological:** Water bacteria, endotoxin, cyanobacteria toxins (anatoxin-A, microcystin-LR)
- **Chemical/Trace Elements:** aluminum, arsenic, barium, cadmium, calcium, chlorine, chloramine, chromium, copper, fluoride, lead, magnesium, mercury, nitrate, potassium, selenium, silver, sodium, sulfate, strontium, zinc
- **Chemical Contaminants Associated With Dialysis Outbreaks:** Aluminum, Chlorine/Chloramine, Calcium and Magnesium, Copper, Strontium

Clinical Effects of Contaminated Water

Symptoms	Chemical Contaminant
Anemia	Aluminum, Chloramines, Copper, Zinc
Bone Disease	Aluminum, Fluoride, Strontium
Hemolysis	Chlorine/Chloramines, Copper, Nitrates
Hypertension	Calcium, Sodium
Hypotension	Bacteria, Endotoxin, Nitrates

Clinical Effects of Contaminated Water

Symptoms	Chemical Contaminant
Metabolic Acidosis	Low pH, sulfates
Muscle Weakness	Calcium, Magnesium
Neurological Deterioration and Encephalopathy	Aluminum
Nausea and Vomiting	Bacteria, Calcium, Copper, Endotoxin, low pH, Magnesium, Nitrates, Sulphates, Zinc

Water Contaminants And The Lowest Level Concentration Associated With Toxicity In Hemodialysis Patients

Contaminant	mg/L
Aluminum	0.06
Chloramines	0.25
Fluoride	1.0
Copper	0.49
Zinc	0.2
Nitrate (as N)	21
Sulfate	200
Calcium/Magnesium	88 (Ca ²⁺)
Sodium	300

Aluminum Intoxications -Anemia, Bone Diseases, and Dementia

Year	Outbreak Description	Cause	Corrective Action
1982	Aluminum intoxication in 27 patients	Exhausted DI tanks	Install Reverse Osmosis and monitor DI tanks
1992	Aluminum intoxication in 27 patients; 3 deaths	Aluminum containing transfer pump (Acid Concentrate)	Discontinue pump use and substitute for a non aluminum containing pump
2007	Elevated serum aluminum levels in 10 patients (16–237 µg/L; median: 92)	Aluminum containing transfer pump (Acid Concentrate)	Discontinue pump use and substitute for a non aluminum containing pump

**FDA SAFETY ALERT:
Aluminum and Other Trace Element Contamination in Dialysis Facilities**

May 20, 1992

TO: HEMODIALYSIS PERSONNEL
WATER OR DIALYSATE SERVICE CONTRACTORS

This is to alert you to a potentially hazardous situation in which dialysis patients have been exposed to dialysate with excessive aluminum levels. These high levels were leached over time from components of the dialysate delivery system. Other trace elements (e.g., iron, copper) could also leach out and contaminate the dialysate in certain circumstances. Please share this Alert with those in your organization who are responsible for water treatment, dialysate delivery system and patient care.


In a recent incident at a large suburban dialysis facility, investigated by the Food and Drug Administration and the Center for Disease Control (CDC), a large number of patients were found to have elevated serum aluminum levels. Three patient deaths were associated with aluminum toxicity.

Preliminary findings indicate that the acidified portion of bicarbonate-based dialysate solution was stored and/or metered to the dialysis patients' proportioning hemodialysis system through an aluminum-containing pump. Aluminum from the pump had leached unexpectedly into the dialysate concentrate during transfer to the patient.

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients

Dr. Matthew Arduino, CDC

A Webber Training Teleclass



Kidney International, Vol. 59 (2001), pp. 746-753

Acute aluminum encephalopathy in a dialysis center caused by a cement mortar water distribution pipe

KENRICK BEREND, GDSBERT VAN DER VOET, and WALTHER H. BOER
Dializ Curaçao, Curaçao, Netherlands Antilles; Toxicology Laboratory, Leiden University Medical Center, Leiden, and Department of Nephrology, University Medical Center Utrecht, Utrecht, The Netherlands

Acute aluminum encephalopathy in a dialysis center caused by a cement mortar water distribution pipe.
Background. In Curaçao, distilled seawater from the water plant was used without further purification for hemodialysis for several decades. A new distribution pipe supplying water to a dialysis center on the island was installed in May 1996. To protect it from corrosion, this pipe was lined on the inside with a cement mortar. Because of the aggressiveness of the distilled water, calcium and aluminum (Al) leached from the cement mortar into the water used to prepare dialysate. This caused a possible hard water syndrome and definite acute Al intoxication.

of liters of dialysate every week, and contaminants present in the water may diffuse across the dialysis membrane and cause intoxications. Therefore, guidelines for the composition of water used to generate dialysate are very strict [1, 2]. To reach this standard, most dialysis centers use more or less elaborate water purification systems, consisting of combinations of water softeners, activated carbon filters, deionizers, and reverse osmosis (RO) [3-6]. We report the combination of an acute aluminum (Al) intoxication and a possible hard water syndrome in a

<http://www.nature.com/ki/journal/v59/n2/full/4492079a.html>

Aluminum Exposure, 2007

- Aluminum intoxication since 1992 is rare and sporadic
- Most aluminum exposure is from ingestion
- Other sources of Aluminum include some granular activated carbons

Aluminum Monitoring in Dialysis patients

- National Kidney Foundation Disease Outcome Quality Initiative (K-DOQI) guidelines recommend serum aluminum testing at least annually in all dialysis patients, and every 3 months in those who receive aluminum-containing medications
 - Cluster in 2007 was detected because of monthly serum aluminum levels
 - routine monitoring of serum aluminum levels can provide a useful tool in preventing serious illness among dialysis patients.

http://www.kidney.org/professionals/KDOQI/guidelines_pedbone/guide14.htm

Fluoride Intoxications, Among US Hemodialysis patients

Year	Description	Cause	Corrective Measures
1980	Fluoride Intoxication in 8 patients, 1 death	Excess fluoride in city water; no water treatment by the dialysis facility	Install pretreatment and Reverse Osmosis unit
1993	Fluoride Intoxication in 9 patients, 3 deaths	Temporary DI water treatment system	DI tanks should be monitored by temperature compensate resistivity meters with both audible and visual alarms

Symptoms of Acute Fluoride Intoxication

- Pruritis
- burning or feverish feeling
- Headache
- nausea or vomiting; or diarrhea
- syncope or near syncope
- pain in the chest, back, or abdomen

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients

Dr. Matthew Arduino, CDC

A Webber Training Teleclass

FDA SAFETY ALERT:
FLUORIDE CONTAMINATION OF HEMODIALYSIS WATER SUPPLY

TO: Hemodialysis Personnel and Water or Dialysate Service Contractors August 19, 1993

This is to alert you to a recent incident in which three hemodialysis patients died and several others were hospitalized after exposure to high levels of fluoride in their dialysate, and to urge that you take certain precautions to prevent other incidents of this kind. Please share this Safety Alert with those within your organization who are responsible for water treatment, dialysate delivery systems (including water treatment systems), and patient care.

In the incident, which was investigated by the Food and Drug Administration (FDA) and the Centers for Disease Control and Prevention (CDC), the fluoride concentration in the dialysate was reported to be 15-25 ppm; the ANSI/AAMI standard identifies 0.2 ppm as the acceptable level of fluoride in dialysate.

The high concentrations occurred because the deionizer used to remove fluoride and other contaminants had become exhausted. Part of the problem may have been the warning lights on the deionizer tanks which are used to alert personnel that the deionizer is exhausted. The previous model at the facility used a single light to indicate that the system was functioning properly; when the light went out, the system was nearing exhaustion and needed replacement or regenerating.

Toxicogenic Cyanobacteria, Brazil 1996

- **Toxins include hepatotoxins and neurotoxins**
 - ◆ Microcystin-LR
 - ◆ Anatoxin A
- **Only known dialysis exposure was in Caruaru, Brazil**
 - ◆ Center received untreated water; water was treated in center with carbon adsorption and DI
 - ◆ 116/130 patients had visual disturbances, nausea and vomiting, and liver failure; at least 50 patients had died
- **If water was treated using reverse osmosis toxins would have been removed**

Jochimsen EM, et al. Liver Failure and Death after Exposure to Microcystins at a Hemodialysis Center in Brazil. *NEJM* 1998;338 (13):873-8
<http://content.nejm.org/cgi/content/abstract/338/13/873>

Human Error and Chemical Intoxications

Patient exposure to disinfectants:

- **Antimicrobial preservatives in filters**
 - ◆ Sodium azide packed ultrafilters
- **Water treatment system disinfectants**
 - ◆ Formaldehyde
 - ◆ Hydrogen peroxide
- **Resizing water distribution system for increased flow without taking into account pre-treatment needs**
 - ◆ Monochloramine exposure
- **Failure to adequately rinse disinfectant from Reprocessed dialyzers**
 - ◆ Peracetic acid
 - ◆ Formaldehyde

Sodium Azide Exposure

- Temporary water treatment system including DI and ultrafilters.
- New industrial ultrafilters installed (not labeled for medical use)
 - ◆ Ultrafilters packed in 0.25% sodium azide and 25% glycerin to prevent bacterial contamination
- Severe life threatening hypotension in 9 patients, other symptoms included blurred vision, severe abdominal pain, headache, and loss of consciousness

Gordon SM, et al. Epidemic hypotension in a dialysis center caused by sodium azide. *KI* 1990; 37:110-115
<http://www.nature.com/ki/journal/v37/n1/pdf/ki199015a.pdf>

Volatile Sulfur Compounds August 30, 2000

- 16 patients developed chills in the absence of fever and hypotension
- Odor (H₂S) in the unit detected
- Other symptoms: nausea, vomiting
- All were hospitalized
- One died within hours


Toxic Effects of Sulfur Containing Compounds

- Toxic effects resulting from inhalation, ingestion or dermal exposure
 - ◆ Gastrointestinal
 - ◆ Respiratory
 - ◆ Central Nervous System
 - ◆ Skin
- NO documented parenteral exposures in humans

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients

Dr. Matthew Arduino, CDC

A Webber Training Teleclass



Volatile Sulfur Compounds Summary

- Outbreak of severe and fatal reactions in hemodialysis patients
- Dialysis water contaminated with sulfur-containing compounds
- Recommendations:
 - ◆ Use pH controller and muriatic acid instead of citric acid
 - ◆ Prompt correction of reverse osmosis unit malfunction
 - ◆ Routine monthly disinfection of the water system

Manufacturing Errors

- Occluded Bloodlines: Acute Hemolysis
- Althin Dialyzers Contaminated with Perfluorocarbon performance fluid

Faulty Blood Tubing sets

Kidney International, Vol. 57 (2000), pp. 1668-1674

Multistate outbreak of hemolysis in hemodialysis patients traced to faulty blood tubing sets

ROSEMARY DUFFY, KAY TOMASHEK, MICHAEL SPANGENBERG, LESLIE SPRY, DIANE DWYER, THOMAS J. SAURANKE, CHRISTOPHER YING, DAVID PORTESI, HOZEFA DIVAN, JOHN KOBRENSKI, MATTHEW ARDUINO, JEROME TOKARS, and WILLIAM JARVIS

Hospital Infection Program and National Center for Infections, Diseases, and International Emergency and Refugee Health Program, National Center for Environmental Health, Public Health Service, U.S. Department of Health and Human Services, Centers for Disease Control and Prevention, Atlanta, Georgia; Office of Regulatory Affairs, Food and Drug Administration, Omaha; and Dialysis Center of Lincoln, Lincoln, Nebraska; Maryland Department of Health and Mental Hygiene, Baltimore, Maryland; Nebraska Health and Human Services System, Lincoln, Nebraska; Lahey Clinic, Burlington, Massachusetts; and Department of Epidemiology, Rollins School of Public Health, Emory University, Atlanta, Georgia, USA

Multistate outbreak of hemolysis in hemodialysis patients traced to faulty blood tubing sets. Background: Hemolysis associated with hemodialysis is rare. The most frequent causes of hemodialysis-associated hemolysis are chemical contamination, heat, or mechanical injury. Conclusions: Our investigation traced the multistate hemolysis with the hemodialysis blood tubing revealed that hemolysis was caused by increased pressure on cylindrical sets as they passed through the partially occluded hemodialysis blood tubing.

<http://www.nature.com/ki/journal/v57/n4/full/4491507a.html>

Partially Occluded Blood Lines

- Patients in three states developed hemolysis while produced by a single manufacturer.
- 35 case patients
- 2 implicated tubing sets of lot 04D15309
- Gambro Healthcare estimated that the degree of occlusion in the defective cassette and tubing varied between 20 and 80%.
- The 300 defective cartridges went to the production of Cobe Centrysystem 3 blood tubing sets of lot numbers 04D15309, 04D15308, and 04D15310.
- Lots were recalled

Fig. 1. Schematic diagram of hemodialysis cassette and blood tubing set. The cassette, circuit, and tubing are from Cobe Centrysystem 3 blood tubing sets of lot numbers 04D15309, 04D15308, and 04D15310. Reprinted with permission from the Centers for Disease Control and Prevention, Atlanta, GA, USA.

Baxter/Althin Hemodialyzers

- In 2001, Baxter announced a voluntary recall of its AX, AF, and A series dialyzers following reports worldwide associated with the use of these dialyzers.
- The models of dialyzers recalled, labeled either Baxter or Althane are: Series A11, A15, A18, A22; Series AF150, AF180, AF220; and Series AX1500, AX2200
- Patients did not respond to any resuscitation efforts

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients

Dr. Matthew Arduino, CDC

A Webber Training Teleclass

Baxter/Althin Hemodialyzers

- The total death count as a result of the use of Baxter's affected dialyzers globally is over 50 hemodialysis patients. Spain, which had the first cases reported, had a total of 12 deaths.
- Highest mortality occurred in Croatia, which had 23 deaths.
- Four deaths occurred in the U.S. with two in Austin, TX, and two in Kearney, NE.
- Deaths also occurred in Colombia, Italy, Germany, and Taiwan

Perfluorocarbon Performace fluid

- PF5070, perfluorocarbons play a prominent role in the process known as "dialysis repair" and has been used throughout the dialysis industry for more than 30 years without a problem
- To repair fibers in the manufacturing process that fail the initial leak test. PF5070 is then allowed to evaporate out of the dialyzer
- PF5070 Characteristics:
 - ◆ liquid at room temperature
 - ◆ a gas at body temperature
 - ◆ insoluble in plasma

For additional information See:
http://biomed.brown.edu/Courses/BI108/BI108_2007_Groups/group05/pages/baxter.html

PF5070 Exposure

- Cardiac arrest
- Respiratory Failure
- Severe hypotension, loss of consciousness
- Dyspnea
- Chest pain/Abdominal pain
- Nausea and vomiting

Gasparovic V, Ostojic R. *Unexpected hemodialysis-related deaths in Croatia. J NEPHROL* 2002; 15: 194-197
<http://www.sin-italy.org/vecchisito/jonline/Vol15N2/194.html>

Allergic Reactions

- May or may not occur in clusters
- Usually patients are easily identified and respond to changes in therapy, ie different reuse chemicals, different dialyzer membrane, preprocessing of hemodialyzers before use, benadryl

ACE Inhibitors and Reuse

Kidney International, Vol. 41 (1992), pp. 1222-1237

Anaphylactoid reactions associated with reuse of hollow-fiber hemodialyzers and ACE inhibitors

DAVID A. PEGUES, CONSUELO M. BECK-SAGUE, STAN W. WOOLLEN, BONNIE GREENSPAN, SUSAN M. BURNS, LEE A. BLANDS, MATTHEW J. ARDUINO, MARTIN S. FAVERO, ROBERT C. MACKOW, and WILLIAM R. JARVIS

Minghai Infectious Program, National Center for Infectious Diseases, Centers for Disease Control, Public Health Service, U.S. Department of Health and Human Services, Atlanta, Georgia, and Food and Drug Administration, Public Health Service, U.S. Department of Health and Human Services, Fairfax, and Contracted District of Springfield Fairfax, Springfield, Virginia, USA

Anaphylactoid reactions associated with reuse of hollow-fiber hemodialyzers and ACE inhibitors. From July 18 through November 27, 1990, 12 anaphylactoid reactions (ARs) occurred in 10 patients at a hemodialysis center in Virginia. One patient required hospitalization, no patient died. ARs occurred within minutes of starting dialysis and were associated with generalized hives and angioedema, respiratory edema or angioedema, facial or generalized constriction of airways, labial numbness or swelling, and 12 ARs occurred with dialyzers that had been reprocessed with an automated reprocessing system. A cohort study, including all patients undergoing dialysis sessions on the six days when an AR occurred, showed that the patients who experienced ARs were significantly more likely than patients who did not to be treated with dialyzers reprocessed with an automated reprocessing system (OR = 7.9; 95% confidence interval = 2.5 to 25.2) and to have been exposed to reused dialyzers rather than to disposable dialyzers (OR = 8.71; P = 0.006). In those sessions using a reused dialyzer, the mean number of dialysis sessions per case was significantly higher than for sessions-reuse (0.5 vs. 6.2; P = 0.004). After reuse of dialyzers was discontinued at the center, no further ARs occurred despite the

fact that none of these reactions were associated with reused dialyzers [7]. In the United States, reuse of disposable hemodialyzers for the same patient in a common practice. From 1977 through 1989, the proportion of centers that reused dialyzers for the same patient increased from 18% to 68% [6]. The primary reasons for dialyzer reuse are the cost savings of reuse and avoidance of first-use syndrome. Reuse procedures are considered safe when standards of the Association for the Advancement of Medical Instrumentation are maintained [8, 9]. Reused dialyzers have been associated with outbreaks of infection with a variety of microorganisms and with an increased incidence of pyrogenic reaction compared with first-use dialyzers [6, 10-12]. In general, these outbreaks resulted from inadequate reprocessing procedures. However, outbreaks of ARs associated with reused dialyzers previously had not been

<http://www.nature.com/ki/journal/v42/n5/abs/ki1992409a.html>

Contaminated Heparin



Acute Allergic-Type Reactions Among Patients Undergoing Hemodialysis --- Multiple States, 2007--2008

On February 1, this report was posted as an MMWR Early Release on the MMWR website (<http://www.cdc.gov/mmwr>).

CDC is investigating an outbreak of acute allergic-type reactions among patients who have undergone hemodialysis since November 19, 2007. The majority of reactions have occurred among adult hemodialysis patients, with onset within minutes of starting a hemodialysis session. Although the cause of the outbreak is unknown and remains under investigation, the majority of reactions occurred in patients who received intravenous heparin produced by Baxter Healthcare Corporation (Deerfield, Illinois). Baxter voluntarily recalled one lot of heparin multichloride vials after learning of these adverse events among patients who received heparin during dialysis. This report describes the ongoing investigation.

CDC was first notified on January 7, 2008, by the Missouri Department of Health and Senior Services (MDHSS) of allergic-type reactions among pediatric hemodialysis patients that occurred beginning November 19, 2007, at a pediatric hospital. The reactions had been reported to MDHSS by a health-care provider at the hospital. The responses occurred within minutes of dialysis initiation and included facial redness, periorbital, perioral, and neck edema. A total of eight episodes of acute allergic-type reactions have been identified as occurring among four patients at the pediatric hospital during November 19, 2007--January 15, 2008. These reactions were resolved by a clinical allergist and were determined to be consistent with anaphylactoid or anaphylactoid reactions.

Upon learning of the initial cluster, CDC solicited reports of similar allergic-type reactions among hemodialysis patients nationally through telephone e-mail lists and public health notifications. In response to these case-finding measures, CDC was contacted on January 9, 2008, by a dialysis supply company that had received reports during the previous 2-week period of approximately 50 similar reactions among adult hemodialysis patients at dialysis facilities in six states. A second supply

Hosted by Paul Webber paul@webbertraining.com
www.webbertraining.com

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients
Dr. Matthew Arduino, CDC
A Webber Training Teleclass

Heparin Related Symptoms

- confirmed case episode of anaphylactic or anaphylactoid reaction characterized by angioedema (particularly swelling of lips/mouth, tongue, throat, or eyelids) or urticaria.
- A probable case had at least two of the following signs and symptoms:
 - ◆ generalized or localized sensations of warmth
 - ◆ numbness or tingling of the extremities
 - ◆ difficulty swallowing
 - ◆ shortness of breath, audible wheezing, or chest tightness
 - ◆ low blood pressure/tachycardia
 - ◆ nausea or vomiting.

Acute Allergic Reactions from Contaminated Heparin

- Oversulfated chondroitin sulfate (OSCS) identified as a contaminant
- Directly activates complement and kallikrein systems
- Contaminated heparin products have now been found in at least 10 countries
- May also stimulate cytokine production
- Products from multiple suppliers received contaminated active pharmaceutical ingredient (API)
 - ◆ Baxter Healthcare
 - ◆ BBraun
 - ◆ Covidien
 - ◆ American Health Products

<http://content.nejm.org/cgi/reprint/NEJMoa0803200v2.pdf>

THE NEW ENGLAND JOURNAL OF MEDICINE

ORIGINAL ARTICLE

Contaminated Heparin Associated with Adverse Clinical Events and Activation of the Contact System

Takashi Kei Kishimoto, Ph.D., Karthik Vivasanathan, Ph.D., Tammy Ganguly, Ph.D., Subbiah Elankumaran, Ph.D., Sean Smith, B.S., Kevin Pelzer, Ph.D., Jonathan C. Lansing, Ph.D., Nammalwar Srianganathan, Ph.D., Garlin Zhao, M.D., Zoya Galcheva-Gargova, Ph.D., Ali Al-Hakim, Ph.D., Gregory Scott Bailey, B.S., Blair Fraser, Ph.D., Sudharita Roy, Ph.D., Thomas Rogers-Cotrone, M.S., Lucinda Buhse, Ph.D., Mark Wharry, Ph.D., James Fox, Ph.D., Moheli Nave, Ph.D., Gerald J. Dal Pan, M.D., Zachary Shriver, Ph.D., Robert S. Langer, Sc.D., Ganesh Venkataraman, Ph.D., K. Frank Austen, M.D., Janet Woodcock, M.D., and Ram Sasisekharan, Ph.D.

- Directly activated the kinin-kallikrein pathway in human plasma
- OSCS induced generation of C3a and C5a,

Non-chemical Associated Hemolysis

- Dialysate temperature $\geq 40^{\circ}\text{C}$
 - ◆ Tielemans CL, Herbaut CR, Geurts JO, Dratwa M. Hemolysis and consumption coagulopathy due to overheated dialysate. *Nephron* 1982;30(2):190-1.
 - ◆ Hecht B, Berkman P, Risch ME. Letter: Hemolysis from hot dialysate. *Ann Intern Med*. 1975 Dec;83(6):902-3.
 - ◆ Berkes SL, Kahn SI, Chazan JA, Garella S. Prolonged hemolysis from overheated dialysate. *Ann Intern Med*. 1975 Sep;83(3):363-4.
- Dialysis against distilled water
 - ◆ Pendergrast JM, Hladunewich MA, Richardson RM. Hemolysis due to inadvertent hemodialysis against distilled water: Perils of bedside dialysate preparation. *Crit Care Med* 2006;34(10):2666-73.
- Kinked Blood tubing

Hemolytic Dialysis Events

- Sometimes rare/sporadic events
- Clusters usually represent exposure to chemical agent
 - ◆ Differentiate from all potential causes
 - * Kinking of tubing sets (Make sure using correct tubing set for machine)
 - * Needle burs
 - * Monochloramine/chlorine exposure
 - * Drug reaction

Bleeding/Exsanguination Events

- Current CDC investigation to determine risk factors
 - ◆ Exsanguination Deaths among Dialysis Patients: District of Columbia, Maryland, Virginia 1/2000 - 7/2007 (Ellingson K, Lucero C, Kurkjian K, Palekar R, Chai D, Schlossberg D)
 - ◆ Access failures
 - * Fistulas, Grafts, catheters
 - ◆ Line separation

Hosted by Paul Webber paul@webbertraining.com
www.webbertraining.com

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients

Dr. Matthew Arduino, CDC

A Webber Training Teleclass

Motivation for Study

- Maryland Medical Examiner review of cases
 - ◆ Bleeding or “exsanguination” deaths in dialysis pts
 - ◆ 24 deaths over 6 years via retrospective review
 - ◆ Since 7/2006 13 additional cases noted by ME
- Little known about epidemiology
 - ◆ Cluster in Maryland?
 - ◆ Incidence in the US?
 - ◆ Case finding?
 - ◆ Preventable risk factors?



CMS Data

DEPARTMENT OF HEALTH AND HUMAN SERVICES
CENTERS FOR MEDICARE & MEDICAID SERVICES

Form Approved
OMB No. 0938-0045

ESRD DEATH NOTIFICATION
END STAGE RENAL DISEASE MEDICAL INFORMATION SYSTEM

1. Patient's Last Name: First _____ MI _____ 2. Medicare Claim Number _____

3. Patient's Sex: a. Male b. Female 4. Date of Birth: / / 5. Social Security Number _____

6. Patient's State of Residence: a. Inpatient b. Home c. Other 7. Place of Death: a. Hospital b. Other Inst. c. Home 8. Date of Death: / /

9. Modality at Time of Death: a. Incenter Hemodialysis b. Home Hemodialysis c. CAPD d. CCPD e. Transplant f. Other

10. Provider Name and Address (Street) _____ 11. Provider Number _____

Provider Address (City/State) _____

- Providers must submit to CMS within 45 days
- Can select up to 5 causes of death
 - ◆ “39” Hemorrhage from Vascular Access
 - ◆ “40” Hemorrhage of Dialysis Circuit

Analysis of CMS Data

- CMS provided CDC with COD data for 2000-07
- Hemorrhage from Vascular access (HVA) and hemorrhage of dialysis circuit (HDC) accounted for **3.7/1000** ESRD deaths with known causes
 - ◆ **5.7/1000** ESRD deaths in MD (6th in US)
 - ◆ **6.5/1000** ESRD deaths in DC (2nd in US)
 - ◆ **5.4/1000** ESRD deaths in VA (7th in US)
 - 1700+ deaths nationwide HVA/HDC codes
 - 18% ESRD deaths listed not coded or unknown


Proportional Death from HVA/HDC Top 10 States



Case Description (N=96)

		N (%) Median (range)
Sex	Female	47 (49.0%)
	Age	64 (28-92)
Race	Black	68 (72%)
	White	24 (25.3%)
	Other	3 (3.2%)

Vascular Access Type (N=96)



Hosted by Paul Webber paul@webbertraining.com
www.webbertraining.com

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients

Dr. Matthew Arduino, CDC

A Webber Training Teleclass

Bleeding and Access Complications (N=73)


- 20.5% had documented history of a serious vascular access hemorrhage
- 72.6% had documentation of ANY access-related complications within the past 6 months
 - ◆ 30.1% Clotting
 - ◆ 27.4% Prolonged bleeding after dialysis
 - ◆ 20.6% Stenosis
 - ◆ 16.4% Superficial infection at access site
 - ◆ 13.7% Extensive vascular access infection
 - ◆ 11.0% Aneurysm or pseudoaneurysm
 - ◆ 4.1% Graft erosion

Summary of Prototypes

- History of access-related problems/concerns
 - ◆ E.g. Infection, prolonged bleeding, erosions, aneurysms, clotting, repairs/revisions
- Anticoagulation-related events
 - ◆ Dialysis heparinization, systemic anticoagulation
- Psychosocial concerns
 - ◆ Depression, mental conditions, financial concerns, substance use, non-compliance
- Medical errors
 - ◆ CVC insertion events, needle dislodgements

Regional Progress on Fistula First

Figure 6 PERCENT OF PREVALENT PATIENTS WITH AV FISTULAS Network 5 Compared to the U.S. - December 2006



Preventing Adverse Events

- Follow standards and recommended practices
- Facility and System Designs
- Quality Assurance Performance Improvement (QAPI)
 - ◆ Surveillance Systems
 - ◆ Data Analysis
 - ◆ Inventory Control
- Documentation
- Know when to ask for help

Water Treatment Systems



AAMI Standards and Recommended Practices

- RD62: Water treatment equipment for hemodialysis applications. ANSI/AAMI RD62-2006.
- Aimed at manufacturer's
- RD52: Recommended Practice-Dialysate for hemodialysis. ANSI/AAMI RD52-2004/A1-2007/A2/2007
 - ◆ Amendment 1- Annex C: Special considerations for home hemodialysis
 - ◆ Amendment 2 - Annex D: Self-assessment of compliance with recommendations for dialysate preparation

Hosted by Paul Webber paul@webbertraining.com
www.webbertraining.com

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients
Dr. Matthew Arduino, CDC
A Webber Training Teleclass

Water for Hemodialysis Applications

- **Defined Chemical agents**
 - ◆ **Group I: Agents known to cause adverse effects in patients**
 - ◆ **Group II: Agents known to be toxic for humans when present in potable water**
 - ◆ **Group III: Chemicals not normally harmful and are present in physiologic fluids and potentially dangerous if present in abnormal concentrations**
- **Defined levels for microbial contamination**

Group 1 Maximum Allowable Chemical Contaminants Water for Hemodialysis Applications

<u>Chemical Agent</u>	<u>mg/L</u>
Aluminum	0.01
Chlorine (free)	0.50
Chloramine	0.10
Copper	0.10
Fluoride	0.20
Nitrate	2.00
Sulfate	100.00
Zinc	0.10

Group II Chemical Contaminant Levels Water for Hemodialysis Applications

<u>Trace Element</u>	<u>mg/L</u>
Antimony	0.006
Arsenic	0.005
Barium	0.10
Beryllium	0.0004
Cadmium	0.001
Lead	0.005
Mercury	0.0002
Silver	0.005
Thallium	0.002
Chromium*	0.014
Selenium	0.09

* Based on the "no-transfer" level

Group III Chemicals: Water for Hemodialysis Applications

<u>Chemical</u>	<u>mg/L</u>	<u>mEq/L</u>
Calcium	2	0.1
Magnesium	4	0.3
Sodium	70	3
Potassium	8	0.2

- Surveillance**
- **On-line monitoring of water quality**
 - Test for chlorine/chloramine prior to each patient shift
 - Test for hardness twice a day
 - On-line monitoring of TDS/Resistivity with temperature compensated meters (audio and visual alarms)
 - **Routine Environmental Cultures of Hemodialysis Fluids (monthly)**
 - **At least annual chemical testing of water (preferably with change of seasons)**
 - **Patient Monitoring**
 - Pyrogenic reactions and/or bacteremia
 - Other adverse patient reactions during dialysis

- Surveillance System**
- **Documentation**
 - ◆ Do you track lot numbers (drugs administered, dialysate concentrate, blood tubing sets, dialyzers, etc)
 - **Separate Log**
 - ◆ Blood stream infections
 - ◆ Hepatitis sero-conversions
 - ◆ Adverse events (note symptoms and circumstances)

Preventing Adverse Outcomes in Maintenance Hemodialysis Patients
Dr. Matthew Arduino, CDC
A Webber Training Teleclass

Understanding a Voluntary Recall

- FDA can encourage the firm to voluntarily correct the problem or to recall a faulty product from the market.

Email Listservers

- Nephrol
 - Renalpro
- To subscribe to Nephrol, RenalPro or other nephrology listserve see <http://www.cybernephrology.org/communication/commProviders.htm>
- DHQP Rapid Notification for Healthcare Professionals
 - http://www2a.cdc.gov/ncidod/hip/rns/hip_ms_subscribe.html
 - Free MMWR Subscription:
 - <http://www.cdc.gov/mmwr/mmwrsubscribe.html>

Additional On-line Resources

CDC Dialysis Pages: www.cdc.gov/ncidod/dhqp/dpac.html
FDA Medwatch and E-listserv: www.fda.gov/medwatch
National Kidney Foundation-Kidney Dialysis Outcomes Quality Initiative (K-DOQI): www.kidney.org/professionals/KDOQI
Kidney Disease: Improving Global Outcomes (KDIGO): www.kdigo.org
United States Renal DataSystems: www.usrds.org
Fistula First: www.fistulafirst.org

THE NEXT FEW TELECLASSES

22 May 08	<i>Bedroom Decontamination - Manual vs. Mechanical</i> Speaker: Gerjo van Klingenberg Gordelink, International Consultant Infection Prevention and Hygiene, The Netherlands
19 Jun. 08	<i>Environmental Sampling - Methods and Strategies</i> Speaker: Dr. Lynne Schuster, CDC
25 Jun. 08	<i>(South Pacific Teleclass) Peripheral Line Sepsis</i> Speaker: Dr. Steve McBride, Auckland District Health Board
26 Jun. 08	<i>CBIC Teleclass 3 - The CIC Examination Process: Computer Based Testing</i> Speaker: CBIC Board Members & Guests
17 Jul. 08	<i>(Free Teleclass) Community-Associated MRSA - What's Up & What's Next</i> Speaker: Dr. Rachel Gorwitz, CDC
22 Jul. 08	<i>(Free British Teleclass) Progress Report from the Chief Nursing Officer</i> Speaker: Christine Beasley, British Department of Health
24 Jul. 08	<i>(Free Teleclass) Disinfection & Sterilization - Current Issues & New Research</i> Speaker: Dr. William Rutala, University of North Carolina

www.webbertraining.com/schedule1.php

Hosted by Paul Webber paul@webbertraining.com
www.webbertraining.com